

# LIGHT SULFUR COMPOUNDS ANALYSIS AT LOW LEVEL

**Technique :** Micro-GC

**Column :** PoraPLOT

**Carrier gas :** Helium

**Column temperature :** 84°C

**Column pressure :** 27.2 psi

**Injection time :** 150 ms

1. H <sub>2</sub> S	50 ppm <sub>v</sub>
2. COS	100 ppm <sub>v</sub>
3. SO <sub>2</sub>	100 ppm <sub>v</sub>

